

# Large Evaluation of Anti-Cardiolipin and anti- $\beta$ 2 Glycoprotein I Assays: Results from the Autoimmunity Workshop of the Spanish Society of Immunology

Marcos López-Hoyos<sup>1</sup>, Dora Pascual-Salcedo<sup>2</sup>, Lourdes Mozo<sup>3</sup>, Maria Rosa Juliá<sup>4</sup>, Margarita Rodríguez-Mahou<sup>5</sup>, Delia Almeida<sup>6</sup>, Montserrat Alsina<sup>7</sup>, Maria Luisa Vargas<sup>8</sup>, Julia Sequí<sup>9</sup>, Raquel Sáez<sup>10</sup>, Maria José Rodrigo<sup>11</sup>, Ángela Carrasco<sup>12</sup>, Pilar Palomino<sup>13</sup>, Jordi Bas<sup>14</sup>, Aresio Plaza<sup>15</sup>, Yvelise Barrios<sup>16</sup>, Alvaro Prada<sup>17</sup>, Rosa Castro<sup>18</sup>, Rosa Maria Pastor<sup>19</sup>, José Marcos García-Pacheco<sup>20</sup>, Luis Fernández-Pereira<sup>21</sup>, Alfonso Sánchez-Ibarrola<sup>22</sup>, Carmen Rodríguez<sup>23</sup>, Mercedes Nocito<sup>24</sup>, Maria José Amengual<sup>25</sup>, Inmaculada Alarcón<sup>26</sup>, Maria Ángeles Figueredo<sup>27</sup>, Laura Jaímez<sup>28</sup>, Juana Jiménez<sup>29</sup>, Antonio Serrano<sup>30</sup>, Esther Ocaña<sup>31</sup>, Eva Martínez-Cáceres<sup>32</sup>, Manuel Santamaría<sup>33</sup>, Neifred Villegas<sup>1</sup>

<sup>1</sup>Hospital Universitario Marqués Valdecilla, Santander. <sup>2</sup>Hospital Universitario La Paz, Madrid. <sup>3</sup>Hospital Universitario Central de Asturias, Oviedo. <sup>4</sup>Hospital Universitario Son Dureta, Palma de Mallorca. <sup>5</sup>Hospital Universitario Gregorio Marañón, Madrid. <sup>6</sup>Hospital Universitario Virgen de la Candelaria, Santa Cruz de Tenerife. <sup>7</sup>Hospital Mútua de Terrassa, Terrassa, Barcelona. <sup>8</sup>Hospital Infanta Cristina, Badajoz. <sup>9</sup>Instituto de Salud Carlos III, Madrid. <sup>10</sup>Hospital Donostia, San Sebastián. <sup>11</sup>Hospital Universitario Vall d'Hebron, Barcelona. <sup>12</sup>Hospital Universitario Ramón y Cajal, Madrid. <sup>13</sup>Fundación Jiménez Díaz, Madrid. <sup>14</sup>Hospital de Bellvitge, L'Hospitalet de Llobregat, Barcelona. <sup>15</sup>Hospital Universitario Puerta de Hierro, Majadahonda, Madrid. <sup>16</sup>Hospital Universitario de Canarias, La Laguna, Tenerife. <sup>17</sup>Hospital de Cruces, Baracaldo, Vizcaya. <sup>18</sup>Laboratorio de Referencia de Catalunya, El Prat de Llobregat, Barcelona. <sup>19</sup>Hospital Joan XXIII, Tarragona. <sup>20</sup>Hospital Universitario Virgen de la Arrixaca, Murcia. <sup>21</sup>Hospital San Pedro de Alcántara, Cáceres. <sup>22</sup>Clínica Universitaria de Navarra, Pamplona. <sup>23</sup>Hospital Universitario Puerta del Mar, Cádiz. <sup>24</sup>Hospital Clínico Universitario de Valladolid, Valladolid. <sup>25</sup>Corporació Sanitària Parc Taulí, Sabadell, Barcelona. <sup>26</sup>Hospital Universitario Dr. Negrín, Las Palmas de Gran Canaria. <sup>27</sup>Hospital Clínico San Carlos, Madrid. <sup>28</sup>Hospital Universitario Virgen de las Nieves, Granada. <sup>29</sup>Hospital Severo Ochoa, Leganés, Madrid. <sup>30</sup>Hospital Universitario 12 de Octubre, Madrid. <sup>31</sup>Complejo Hospitalario de Jaén, Jaén. <sup>32</sup>Hospital Universitario Germans Trias i Pujol, Badalona, Barcelona. <sup>33</sup>Hospital Universitario Reina Sofía, Córdoba.

## EVALUACIÓN DE ENSAYOS ANTI-CARDIOLIPINA Y ANTI- $\beta$ 2-GLICOPROTEÍNA I: RESULTADOS DEL TALLER DE AUTOINMUNIDAD DE LA SOCIEDAD ESPAÑOLA DE INMUNOLOGÍA

Recibido: 21 Abril 2009

Aceptado: 10 Junio 2009

### RESUMEN

A pesar de la indudable utilidad clínica y de la importancia de las pruebas de laboratorio en el diagnóstico del síndrome antifosfolípido (APS), probablemente el mayor defecto de dichas pruebas es su elevada variabilidad intra- e inter-laboratorio. El objetivo del presente trabajo fue evaluar el comportamiento de los ensayos para detección de anticuerpos anti-cardiolipina (aCL) y anti-beta 2 glicoproteína I (anti- $\beta$ 2GPI) entre laboratorios y determinar el grado de variabilidad inter-laboratorio e inter-ensayo. En este trabajo se describen los resultados más significativos del Taller de Autoinmunidad de la SEI. 17 sueros obtenidos de pacientes con APS y/o probable APS se recogieron tras consentimiento informado. 33 laboratorios participaron y midieron los títulos de aCL y anti- $\beta$ 2GPI. 61 y 49 resultados/suero se informaron para aCL and anti- $\beta$ 2GPI (IgG/IgM), respectivamente, y medidos con 20 ensayos diferentes. Se encontró un coeficiente de variación (CV) elevado en los resultados cuantitativos, inde-

pendientemente del método empleado. El CV fue del 50-128% para aCL y 9-200% para anti- $\beta$ 2GPI. Se obtuvo un consenso (definido como >90% de acuerdo) débil para los resultados semicuantitativos de IgG/IgM aCL y anti- $\beta$ 2GPI: 47%, 65%, 47% y 70%, respectivamente. En general, hubo una buena concordancia entre aCL y anti- $\beta$ 2GPI, aunque 2 de los 17 sueros fueron positivos para anti- $\beta$ 2GPI pero no para aCL. En resumen, la interpretación de los resultados de aCL y anti- $\beta$ 2GPI emitidos por distintos laboratorios puede hacerse solo en términos semicuantitativos y su valor real en el diagnóstico clínico del APS es aún limitada. Los puntos de corte para cada ensayo deben ser establecidos por el propio laboratorio.

PALABRAS CLAVE: Síndrome Antifosfolípido/ SAF/ Anticuerpos antifosfolípido/ Anticoagulante de Lupus/ Anti-cardiolipina/ anti- $\beta$ 2-glicoproteína I.

**ABSTRACT**

Despite their clinical utility and the importance that laboratory tests have in APS diagnosis, probably the most important drawback of such tests is the elevated intra- and inter-laboratory variation. The aim of the present work was to assess the multilaboratory performance of aCL and anti- $\beta$  the multilaboratory performance of anti-cardiolipin (aCL) and anti-beta 2 glycoprotein I antibodies (anti- $\beta$ 2GPI) assays and to assess the inter-laboratory and inter-assay variability. Here, we report the most significant results from the Autoimmunity Workshop of the Spanish Society of Immunology (AWSEI). Seventeen sera from patients with antiphospholipid syndrome (APS) and/or probable APS were collected after written informed consent. Thirty-three laboratories participated and measured aCL and anti-anti- $\beta$ 2GPI. 61 and 49 results/serum for IgG/IgM aCL and anti-anti- $\beta$ 2GPI, respectively, were informed with 20 different assays. A high interlaboratory variation was found in quantitative results regardless the method used. Coefficient of variation ranged from 50% to 128% for aCL and from 9% to 200% for anti-anti- $\beta$ 2GPI. A limited consensus (defined as >90% agreement) was observed in semiquantitative results for IgG/IgM aCL and anti- $\beta$ 2GPI: 47%, 65%, 47% and 70%, respectively. In general, there was concordance between aCL and anti- $\beta$ 2GPI, yet 2 of the 17 sera were positive for anti- $\beta$ 2GPI only. In conclusion, interpretation of aCL and anti- $\beta$ 2GPI results from different laboratories may be done only in semiquantitative terms and its real value for clinical diagnosis of APS is still limited. Cut off values must be set in each laboratory.

**KEY WORDS:** Antiphospholipid Syndrome/ APL/ Anti-phospholipid antibodies/ Lupus anticoagulant/ Anti-cardiolipin antibodies/ Anti- $\beta$ 2-glycoprotein I antibodies.

**INTRODUCTION**

The presence of anti-phospholipid antibodies (APL) is mandatory in the diagnosis of antiphospholipid syndrome (APS)<sup>(1)</sup>. The three clinically useful APL are lupus anticoagulant (LA), anti-cardiolipin antibodies (aCL) and anti-beta 2-glycoprotein I antibodies (anti- $\beta$ 2GPI)<sup>(2)</sup>. Their utility is partially hampered by the deficient standardization of the assays to measure them<sup>(3)</sup>. Whereas LA consists on several coagulation tests, aCL and anti- $\beta$ 2GPI are usually measured by ELISA that, in the beginning, used to be home-made. However, the importance of these tests, as judged by the number of orders, is exponentially increasing in the routine at the same time as APS suspicion from the clinicians is growing. This results in a high pressure on clinical laboratories, forcing them to use commercial kits instead of procedures developed in-house<sup>(4)</sup>. These commercial assays are highly variable and with very different sensitivity/specificity. Altogether, the standardization of the results for aCL and anti- $\beta$ 2GPI antibodies remains a daily problem in the autoimmunity laboratories that perform these important assays for APS diagnosis.

The aim of the Autoimmunity Workshop of the Spanish Society of Immunology (AWSEI) was to evaluate the multilaboratory performance of aCL and anti- $\beta$ 2GPI assays and to assess inter-laboratory and inter-assay variability. Here, we report the most significant results from the workshop.

**MATERIAL AND METHODS****Participants**

The report includes the test data results submitted by 33 laboratories from the 43 centers all over Spain that participated in the AWSEI.

**Serum samples and patients**

Seventeen sera from patients with APS and/or probable APS (Table I) were collected after written informed consent, sent to the coordinator laboratory, aliquoted, and frozen at -80°C until distribution. Frozen aliquots were distributed among participating laboratories in only one package. Aliquots contained serum volume (200-500  $\mu$ l) enough to perform multiple assays (and not only those routinely done in each laboratory).

**Assays and collection of results**

Manufacturers were kindly invited to provide their ELISA kits to the participant laboratories. Table II shows the listing of the manufacturers that contributed with their kits to the AWSEI.

Participating laboratories were asked to give the results as quantitative and semiquantitative (negative, low-, medium- or high-positive titer). When quantitative results were reported as greater than a given number, the numerical value was taken as the next highest whole number. Variation in numerical results between laboratories was determined using the calculated coefficient of variation (%CV), by dividing the standard deviation obtained for the pooled numerical results of the sample by the mean numerical value for that serum. Variations in both quantitative and semiquantitative results were calculated from all reported results and separately for each manufacturer when the number of results reported was  $\geq 5$ .

**RESULTS****The quantitative results show a wide interlaboratory variability**

The quantitative results reported from the participant laboratories for aCL and anti- $\beta$ 2GPI were pooled to determine the inter-laboratory variability regardless of the method used (Table III). Globally, the results showed a very poor agreement between laboratories when using quantitative data to report APL results. The range of % CV was 48.02-115.00, 51-147.1, 49.34-110.35, 46.29-100.54 for aCL IgG, aCL IgM, anti- $\beta$ 2GPI IgG, and anti- $\beta$ 2GPI IgM, respectively. Table III shows that the higher % CV were obtained when antibody titers were

**TABLE I.** Clinical and demographic characteristics of patients whose sera were studied for APL in the workshop

Serum	Gender (M/F)	Age (ys)	Primary Diagnosis	Manifestations	Current treatment
1	F	26	PAPS	Ischemic neuritis; thrombocytopenia	Steroids
2	F	46	Rhupus	Arthritis; cryoglobulinemia, lupus anticoagulant	Steroids, hydroxychloroquine, metotrexate
3	F	61	Probable Conectivopathy	DVT; hypertension, hyperlypemia	Unknown
4	F	23	PAPS	3 FL (wk 20-24); repetitive DVT	Warfarin, hydroxychloroquine
5	F	59	PAPS	2 DVT; 1 TEP; 1 cerebral stroke; 1 abortion	Warfarin
6	F	28	SLE	Unactive SLE, livedo reticularis	Steroids, AZA, hydroxychloroquine
7	F	54	PAPS	Leriche syndrome, renal artery stenosis, smoking	Unknown
8	M	29	TEP	PE; flebitis superficial	Warfarin
9	M	58	SLE	Lupus nephritis, SAPS	Steroids, cyclophosphamide
10	F	20	PAPS	PE	Warfarin
11	M	U	No disease	Brother of patient 8	None
12	F	40	SLE	Arthritis; synovitis; b2PI positive repetitively	U
13	M	45	PAPS	DVT 6 ys ago; ANA and anti-dsDNA positive	U
14	M	43	Uncomplete SLE	DVT	Steroids, hydroxychloroquine, warfarin
15	U	U	U	U	U
16					
17	M	63	Behcet	SAPS	Warfarin

AZA: Azathyoprine; DVT: Deep venous thrombosis; PAPS: Primary antiphospholipid syndrome; PE: Pulmonary embolism; SAPS: Secondary antiphospholipid syndrome; U: Unknown.

low. The interlaboratory variability remained high even when results were grouped by manufacturer (not shown).

#### Report of semiquantitative results demonstrated a moderate agreement between laboratories

Figure 1 shows the results reported for each of the 17 sera as semiquantitative results (i.e., negative, low-, medium- or high-positive titres). To determine the consensus between laboratories in reporting the results, a >90% agreement was established for each antibody and each serum. Thus, negative and low-positive results were considered as a negative result whereas medium- and high-positive results were considered as positive results, as recommended by international consensus<sup>(1)</sup>. A limited consensus (defined as ≥90% agreement) for semiquantitative results was observed: aCL IgG (47%), aCL IgM (65%), anti-β2GPI IgG (47%), anti-β2GPI IgM (70%).

When looking at the overall semiquantitative results depicted in Figure 1, several findings are observed. First, most sera were both positive for aCL and anti-β2GPI of the same isotype, but for sera 5 and 6. Serum 5 was aCL and anti-β2GPI IgM positive in several laboratories but only aCL IgG positive in more than 50% of the reported results. Serum 6 was aCL IgM positive but not anti-β2GPI IgM.


**TABLE II.** Number of assays reported for aCL and anti-β2GPI

Manufacturer	aCL	anti-β2GPI
Aeskulab	8	9
Chesire	1	
Corgenix	1	
Diasorin	9	9
Elia Phadia	7	1
Eurodiagnostica	1	1
Imtec	2	2
Inova	8	9
Orgentec	15	14
Varelisa Phadia	3	1
In house	4	

Thus, these two sera could be considered as aPL positive cofactor-independent. However, sera 15 and 16 were mostly reported positive only for anti-β2GPI IgG but not aCL IgG. On the other hand, APL IgG (sera 4, 8, 11, 13, and 14) were more represented than APL IgM (sera 2, 5, and 7). Finally, serum 17 from a patient diagnosed with Behcet and APS was reported as negative by almost all the participants.

**TABLE III.** Quantitative results reported for each of the 17 sera studied showing the interlaboratory variation and considering all the methods as a whole

Serum	Mean ± SD				Variation coefficient (%)			
	aCL IgG	aCL IgM	β2GPI IgG	β2GPI IgM	aCL IgG	aCL IgM	β2GPI IgG	β2GPI IgM
1	10.5 ± 7.7	13.4 ± 11.2	8.3 ± 6.0	15.5 ± 11.6	73.7	84.0	72.0	74.6
2	23.4 ± 16.6	163.1 ± 119.2	38.3 ± 29.2	182.8 ± 106.9	71.1	73.1	76.1	58.5
3	6.5 ± 6.9	6.3 ± 7.4	6.0 ± 5.2	13.1 ± 26.5	106.2	118.2	87.2	75.2
4	99.6 ± 47.9	17.8 ± 12.7	75.6 ± 37.3	7.6 ± 5.7	48.0	71.2	49.3	75.2
5	46.7 ± 28.7	137.5 ± 77.2	23.2 ± 21.2	150.3 ± 82.1	61.4	56.2	91.5	54.7
6	21.8 ± 25.2	62.1 ± 91.3	7.5 ± 5.9	10.3 ± 6.6	115.8	147.1	78.1	64.7
7	21.1 ± 14.0	94.9 ± 49.0	15.2 ± 11.3	112.6 ± 71.3	66.6	51.6	74.2	63.3
8	147.6 ± 82.2	11.1 ± 7.8	205.1 ± 161.0	7.8 ± 5.4	55.7	69.8	78.5	69.6
9	15.3 ± 10.3	5.7 ± 6.4	23.0 ± 17.7	5.4 ± 4.9	67.3	113.3	77.0	90.1
10	99.6 ± 57.7	6.8 ± 6.7	41.8 ± 28.8	5.1 ± 5.1	57.9	99.4	61.8	100.6
11	137.3 ± 83.9	13.6 ± 8.0	209.3 ± 158.8	11.0 ± 6.9	61.1	58.8	75.9	62.1
12	18.5 ± 15.5	69.2 ± 56.3	29.2 ± 32.2	163.7 ± 75.8	83.7	81.4	110.4	46.3
13	60.7 ± 38.5	22.1 ± 16.1	66.2 ± 42.0	44.2 ± 27.9	63.5	72.7	63.4	63.1
14	151.6 ± 91.5	5.6 ± 6.5	179.2 ± 161.1	4.9 ± 5.2	60.3	115.0	89.9	104.7
15	7.9 ± 7.1	6.7 ± 6.3	74.3 ± 66.1	10.9 ± 8.5	90.4	94.2	89.0	78.2
16	15.4 ± 11.2	6.2 ± 6.5	56.2 ± 45.0	5.9 ± 4.9	72.9	104.8	80.1	83.4
17	15.2 ± 13.9	8.9 ± 7.2	5.5 ± 5.3	5.5 ± 5.1	91.9	81.6	97.1	92.2



**Figure 1.** Pie graphs showing the semiquantitative results reported by all the participant laboratories. Four results for each serum are depicted: aCL IgG (upper left pie), aCL IgM (upper right pie), anti-β2GPI IgG (lower left pie) and anti-β2GPI IgM (lower right pie). The number in the middle indicates the serum identification. Results were reported as negative, low-, medium- or high-positive as indicated in the lower right corner of the figure. To evaluate the congruence of reports as semiquantitative results, negative and low-positive sera were considered as negative whereas medium- and high-positive sera were considered as positive.

## DISCUSSION

More than 25 years after aCL testing in the diagnosis of APS has led us to a number of changes in laboratory assays as well as in APS management. As a consequence APS diagnosis has become more effective, although there is still an important field for improvement in both clinical management and laboratory tests<sup>(5,6)</sup>. Laboratory testing is based in aCL, LA and, more recently, anti- $\beta$ 2GPI. The present scenario indicates that aCL are the most sensitive autoantibodies for the diagnosis of APS but have a very poor specificity. On the other hand, LA and anti- $\beta$ 2GPI seem to be more specific but not as sensitive as aCL.

Despite their clinical utility and the importance that laboratory tests have in APS diagnosis, probably the most important drawback of such tests is the elevated intra- and inter-laboratory variation<sup>(3,4)</sup>. As a result, a patient classified as APS in a hospital may not be considered as APS in another centre where the APL testing may be very different. Inter-laboratory results are not interchangeable even if measured with the same commercial kit because there is still a high inter-laboratory coefficient of variation (data not shown but obtained in the AWSEI), particularly for quantitative results. Such variability in quantitative results has been assumed for anti- $\beta$ 2GPI results since there is no international calibrator to give results in the same units as occurs for aCL (i.e., GPL or MPL). Nonetheless, the high inter-laboratory variation was observed for both aCL and anti- $\beta$ 2GPI within the same commercial kit in the AWSEI, which does not make sense for the previous explanation. When semiquantitative results were reported (Figure 1), variability remained elevated. A possible explanation for this limited reproducibility might reside in the use of different values to define low-, medium- or high-positive titres among laboratories. International consensus establishes that medium/high titres of APL are levels above 40 GLP or MPL units or higher than the 99th percentile of normal subjects<sup>(2)</sup>. The use of different values among laboratories, without a uniform criterion, may be one of the causes explaining the limited inter-laboratory consensus when reporting the results as semiquantitative. The establishment of cut-off values is still a matter of debate since there could be different cut-off levels depending on the clinical subtype of APS patient considered.

An additional message to draw from the present workshop is the variable spectrum of APL observed. As expected, most sera were positive for aCL and anti- $\beta$ 2GPI at the same time due to the cofactor dependence of aCL. On the other hand, there were two sera (15 and 16 in Figure 1) negative for aCL but positive for anti- $\beta$ 2GPI. In the last years, these autoantibodies are gaining importance as prognostic markers

of the disease, particularly high avidity anti- $\beta$ 2GPI antibodies<sup>(7)</sup>. Finally, one serum (from patient 6 with SLE) was aCL IgG and IgM positive but anti- $\beta$ 2GPI IgG and IgM negative, and serum from patient 10, diagnosed with primary APS, was more frequently reported as medium-high positive for aCL IgG than for  $\beta$ 2GPI IgG. These data bring the message of the importance of reporting the four parameters in the revised criteria established in Sidney for the diagnosis of APS.

In conclusion, despite the necessary use of the APL test for APS diagnosis, there is still a long way to walk to perform these laboratory assays in a more precise and reproducible manner. It is probable that such a desired reproducibility would come with the description of the true epitope/ antigen involved in the APS.

## CONFLICT OF INTEREST

The authors declare no financial conflict of interest.

### CORRESPONDENCIA:

Marcos López-Hoyos  
Servicio Inmunología,  
Hospital Universitario Marqués de Valdecilla  
39008 Santander (Spain)  
Phone: +34 942203453. Fax: +34 942203453  
E-mail: inmlhm@humv.es

## REFERENCES

1. Giannakopoulos B, Passam F, Ioannou Y, Krilis SA. How we diagnose the antiphospholipid syndrome. *Blood* 2009;113:985-994.
2. Miyakis S, Lockshin MD, Atsumi T, Branch DW, Brey RL, Cervera R, et al. International consensus statement on an update of the classification criteria for definite antiphospholipid syndrome (APS). *J Thromb Haemost* 2006;4:295-306.
3. Reber G, Boehlen F, de Moerloose P. Technical aspects in laboratory testing for antiphospholipid antibodies: Is standardization an impossible dream? *Semin Thromb Hemost* 2008;34:340-346.
4. Favaloro EJ, Wong RCW. Laboratory testing and identification of antiphospholipid antibodies and the antiphospholipid syndrome: A potpourri of problems, a compilation of possible solutions. *Semin Thromb Hemost* 2008;34:389-410.
5. Bertolaccini ML, Khamashta MA, Hughes GRV. Diagnosis of antiphospholipid syndrome. *Nat Clin Pract Rheumatol* 2005;1:40-46.
6. Erkan D, Lockshin MD. New approaches for managing antiphospholipid syndrome. *Nat Clin Pract Rheumatol* 2009;5:160-170.
7. Cucnik S, Kveder T, Krizaj I, Rozman B, Bozic B. High avidity anti- $\beta$ 2-glycoprotein I antibodies in patients with antiphospholipid syndrome. *Ann Rheum Dis* 2004;63:1478-1482.