



Review

Videolaryngoscopes: The solution for difficult airway management or just another strategy? Non-systematic review[☆]



Katheryne Chaparro-Mendoza^{a,*}, Carlos Andrés Luna-Montúfar^b, Juan Manuel Gómez^c

^a Cardiovascular and Thoracic Anesthesiologist, Fundación Clínica Valle del Lili, Cali, Colombia

^b Anesthesiologist, Instructor in the Department of Anesthesia, Universidad del Valle, Hospital Universitario Evaristo García, Cali, Colombia

^c Pediatric Anesthesiologist, Instructor in the Department of Anesthesia, Universidad del Valle, Centro Médico Imbanaco, Cali, Colombia

ARTICLE INFO

Article history:

Received 29 July 2014

Accepted 8 March 2015

Available online 18 June 2015

Keywords:

Videotape Recording

Laryngoscopy

Airway Management

Intubation

Intratracheal

Obesity

Anesthesia

Emergencies

ABSTRACT

Introduction: Difficult airway management remains a challenge and is a pillar of anesthesia training. At present, unsuccessful management of the difficult airway is a leading cause of complications in the practice of anesthesia, something that has led to regular updates to the management algorithms and the development of new technologies.

Objectives: To review the current state of videolaryngoscopy and its impact on difficult airway management.

Methods: With the keywords Videotape Recording; Laryngoscopy; Airway Management; Intubation; Intratracheal; Obesity; Anesthesia. A non-systematic review in the following databases was conducted: Pubmed/Medline, SciElo, LILACS).

Results: Videolaryngoscopes are a new technology for the management of difficult airways that so far have not replaced the standard airway management algorithm devices. Its main impact is better visualization of the laryngeal structures. However, there are still controversies regarding the ease and success of tracheal intubation. Evidence of its usefulness in difficult airway management is weak.

Conclusion: Knowledge of these devices and their limitations is an alternative in difficult airway scenario, but its real value and safety for the patient is still not defined and continues to be researched.

© 2015 Sociedad Colombiana de Anestesiología y Reanimación. Published by Elsevier España, S.L.U. All rights reserved.

[☆] Please cite this article as: Chaparro-Mendoza K, Luna-Montúfar CA, Gomez JM. Videolaryngoscopes: ¿la solución para el manejo de la vía aérea difícil o una estrategia más? artículo original. Rev Colomb Anesthesiol. 2015;43:225-233.

* Corresponding author at: Av. Simón Bolívar, carrera 98 N°18-49. Fundación Clínica Valle del Lili, Departamento de Anestesiología y Reanimación, Cali, Colombia.

E-mail address: jacok20@hotmail.com (K. Chaparro-Mendoza).

Videolaringscopios: ¿la solución para el manejo de la vía aérea difícil o una estrategia más? Revisión no sistemática

R E S U M E N

Palabras clave:

Grabación en video
Laringoscopia
Intubación intratraqueal
Vía aérea difícil
Obesidad
Anestesia
Emergencias

Introducción: el manejo de la vía aérea difícil continúa siendo un reto y es uno de los pilares del entrenamiento en anestesia. En la actualidad, el manejo no exitoso de la vía aérea difícil representa una de las principales causas de complicación en el ejercicio de la anestesia que promueve la actualización regular de los algoritmos de manejo y al desarrollo de nuevas tecnologías.

Objetivos: presentar el estado actual de los videolaringscopios y su impacto en el manejo de la vía aérea.

Métodos: con las palabras claves: Grabación en video Laringoscopia, Intubación intratraqueal; Vía aérea difícil; Obesidad; Anestesia; Emergencias se realizó una revisión no sistemática en bases de datos (PubMed/Medline, Scielo, Lilacs).

Resultados: los videolaringscopios son una tecnología adicional para el manejo de la vía aérea que hasta el momento no han demostrado sustituir los dispositivos estándares expuestos en el algoritmo de manejo de la vía aérea. Su principal impacto está determinado por la mejoría en la visualización de las estructuras de la laringe sin embargo aún hay controversias respecto a la facilidad y éxito de la intubación endotraqueal. La evidencia de su utilidad en el manejo exitoso de la vía aérea difícil es débil.

Conclusiones: el conocimiento de estos dispositivos así como sus limitaciones constituye una alternativa en el escenario de la vía aérea difícil, pero su valor real y la seguridad que representa para el paciente aún no se han definido y continúa en investigación.

© 2015 Sociedad Colombiana de Anestesiología y Reanimación. Publicado por Elsevier España, S.L.U. Todos los derechos reservados.

Introduction

Difficult airway is defined as the clinical situation in which a trained anesthesiologist experiences difficulty in ventilation with a face mask or tracheal intubation¹⁻³. Its incidence in the general population is between 1.15 and 3.8%, and that of failed intubation is 0.13–0.3%^{4,5}. The situation may result in complications as severe as bronchoaspiration, lesions in the upper airway, cerebral hypoxia, and death^{1-3,5}.

Awareness of new alternatives for securing the airway is a constant necessity⁵. Video laryngoscopes are a new generation of devices that allow direct visualization of the glottis and have recently been included in several societies' algorithms for airway management. In our context, however, there are few publications about their use, success rate, and safety.

Methodology

A non-systematic literature review in English and Spanish in the databases PubMed/Medline, Scielo, and Lilacs with the following MeSH and DeCS terms: Videotape Recording; Laryngoscopy; Airway Management; Intubation, Intratracheal; Obesity; Anesthesia. We proceeded to read each article and review the relevant references related to videolaringscopios in airway management that allowed us to describe their main characteristics and impact. Finally, 51 articles were chosen through consensus among the three researchers.

Results

The tubular, remote view of the glottis with direct laryngoscopy provides a 15° visual field. This can be extended to between 45° and 60° with videolaringscopios.⁶⁻⁸ The videolaringscope models can be classified according to the mechanism for visualizing the glottis and the design of the blade (Table 1). Depending on the mechanism for visualizing the glottis, they can be:

1. Devices with a miniature video camera incorporated into the distal part of the laryngoscope blade. From here, the image is transmitted to an external screen. Example: McGrath, Glidescope, Storz, King Vision^{6,7,9}.
2. Devices in which the image is transmitted through a fiber optic bundle or through a system of prisms to a storage device, such as a video system or lens. Examples: Airtraq (lenses and prisms) and Bullard (fiber optics)^{6,7,9}.

They are also distinguished by blade type:

1. Videolaringscopios with standard Macintosh blades are inserted using the same technique as in direct laryngoscopy. Example: Storz^{7,9}.
2. Videolaringscopios with angled blades. They have an extra curve that allows for visualization through the camera only. Examples: Glidescope and McGrath^{7,9}.
3. Videolaringscopios with channel blades. They have a central channel through which the endotracheal tube (ETT)

Table 1 – Videolaryngoscope characteristics.

Videolaryngoscope	Visualization of glottis	Type of blade	Size	Portability	Recommendation	Characteristics
Glidescope MODELS: Original Ranger Cobalt	External monitor Anti-fog mechanism	Angled blade MODELS Original: reusable Ranger: transreflective display for bright environments Can be reusable or single use Cobalt: Blade protector so that it does not come into contact with the patient. Single use.	Original: 2-5 Ranger: reusable 3-4 Single use: 1-4 Cobalt: 1-4	Only the Ranger is portable	Use of hockey stick shaped stylet	Insertion along midline, over the back of the tongue. Reported benefits in difficult airway
King Vision	External monitor Anti-fog mechanism	Angled blade, models with or without channel	Single use	Yes	May require use of hockey stick shaped stylet if no channel	Insertion along midline over the back of the tongue or lateral commissure. Model with channel requires greater oral aperture. Easy detachment of the handle that contains the battery and the screen during manipulation
Storz MODELS: V Mac C Mac	External monitor V Mac: 8" monitor C Mac: 7" monitor with anti-fog mechanism	Standard Macintosh	V Mac: Pediatric and adult C Mac: 2-4	Only C Mac	Use of bougie or stylet	Insertion similar to Macintosh blade. Allows direct visualization of structures (advantage in cases of secretions and teaching). The Videolaryngoscopia is used to lift the jaw and submandibular tissues
McGrath	External monitor 2.5" monitor Anti-fog mechanism	Angled single use blade	Only for adults	Yes	Requires hockey stick shaped stylet	Use similar to Glidescope
Airtraq	Lenses and prisms Attachable monitor	Blade with channel, anti-fog system	Three sizes available	Yes	It is located above the glottis so that the tube is directed toward it	Insertion along the midline over the back of the tongue Allows for intubation in pediatrics
Bullard	Fiber optic Attachable monitor	Blade with channel	Four sizes available	No		Requires insertion along the midline, does not have panoramic view. Has a pediatric model

Source: Authors.



Fig. 1 – Differences in the alignment of axes in the use of direct and indirect laryngoscopes.
Source: Authors.

can be preloaded, which allows for insertion once the glottic opening is viewed. Examples: King Vision, Airtraq, and Bullard^{7,9}.

Insertion technique

The insertion of videolaryngoscopes differs from that of conventional laryngoscopes. The alignment of the oral, pharyngeal, and laryngeal axes is not required (Fig. 1). Achieving an adequate oral aperture is essential, since the device must enter along the midline, following the shape of the palate and the posterior pharynx in a way similar to the insertion of laryngeal masks^{7,10,11}.

A frequent difficulty with the use of these devices arises during the insertion of the ETT due to the lateral right displacement of the tongue or to an inadequate oral aperture.



Fig. 2 – Oral aperture maneuver with the thumbs and index fingers.
Source: Authors.

For this, the jaw-thrust maneuver is recommended with finger pressure on the incisors (Fig. 2)^{7,10,11}.

Even with an adequate visualization of the glottis, the insertion of the ETT may be difficult^{10,11}. For this reason, videolaryngoscopes have been designed that have a channel for the insertion and movement of the tube. Also is recommended the use of stylet with a 60° angle in the distal part of the ETT in a shape similar to that of a hockey stick, entering through the oral commissure, and performing the BURP maneuver (Back Up Right Pressure)^{10,11}. Advancing the ETT may also present difficulties since the angle of incidence between the laryngoscope blade and the trachea may make the tube hit the tracheal cartilage¹².

Glidescope

Currently, there are three models of this type of videolaryngoscope: the original Glidescope, the Glidescope Ranger, and the Glidescope Cobalt¹³⁻¹⁶. The original Glidescope is a reusable plastic device with a handle similar to that of a conventional laryngoscope, and a blade with a 60° angle in the middle and a digital camera in the distal part of the blade¹⁴. The Glidescope Ranger has a portable design with a smaller, 3.5" screen with a transreflective display that allows the operator to view the anatomy in a brightly-lit environment, such as in pre-hospital or military settings¹⁶. The Glidescope Cobalt has a blade similar to the Macintosh blade with a 60° angle in the tip. It has a reusable color video camera with a powerful light source and a transparent plastic disposable blade through which the video baton is inserted so that direct contact between the camera and the patient can be avoided. As such, it does not require disinfection¹⁴.

Studies on Glidescope have shown success rates of more than 94%, with intubation times under one minute, and



Fig. 3 – King Vision videolaryngoscope: insertion.

Source: Authors.

improvement of the view of the vocal cords by one to two degrees^{13,16,17} even in patients with restricted cervical mobility, such as those with ankylosing spondylitis¹⁸.

A meta-analysis that compared endotracheal intubation with Glidescope to intubation with direct laryngoscopy demonstrated an improvement in glottic visualization with Glidescope. This benefit increases in patients with difficult airway. Nevertheless, a greater intubation success rate and lower intubation time was only found with Glidescope when performed by inexperienced personnel. No difference was found compared to direct laryngoscopy when intubation was performed by experienced personnel¹⁹. Alteration of neck anatomy was reported as an important predictor of failure with Glidescope²⁰.

King Vision

The King Vision videolaryngoscope is a device with a 2.4" LED screen (160° panoramic view), a disposable blade, and a video connection. It has two types of blades: a standard blade that allows for free manipulation of the ETT with a 60° angled stylet and that requires a minimum oral aperture of 13 mm and insertion along the midline; and a blade with a channel through which the ETT may be introduced and that requires a minimal oral aperture of 18 mm with insertion along midline or laterally (Fig. 3). If the tube hits the right arytenoid cartilage, rotating the videolaryngoscope to the left until obtaining alignment with the entry of the glottis is recommended. Once the ETT has entered the larynx, the stylet should be partially withdrawn and the tube should be rotated 90° to avoid contact with the tracheal cartilages. It is also possible to insert a bougie and run an ETT through it²¹.

In a study conducted on inexperienced personnel, the King Vision without a channel showed a higher success rate and a longer intubation compared to the device with the channel

and the conventional laryngoscope. Between the latter two, there was no difference²².

In the simulated difficult airway setting, the King Vision had a greater success rate and better glottis visualization compared to the traditional laryngoscope²³.

Storz videolaryngoscope

This videolaryngoscope was designed by Karl Storz who modified the Macintosh blade and the handle from a traditional laryngoscope. It has an 8" camera adapted to the handle, which increases the image of the anatomical structures, and a Macintosh blade containing a light that is directed toward the portion of the larynx to be viewed⁷. Its insertion is similar to that of the traditional laryngoscope, with the possibility of directly viewing structures during the process.

There are two models of this type of videolaryngoscope: the V Mac, that features a camera incorporated into the laryngoscope handle, and the C Mac, the newest model with a better image (Fig. 4) and a memory card⁷.

The Storz videolaryngoscope has proven useful in teaching laryngoscopy as it permits a direct view of the anatomical structures and the results of the external manipulation of the larynx. A study by Storz showed that intubation attempts were successful, with a short learning curve, and greater external manipulation of the larynx when difficult airway predictors were present²⁴. In addition, it improved the visualization of the glottis in as many as 40% of patients²⁵, has a success rate of 93% on the first attempt, and required less external larynx manipulation and bougie use compared to direct laryngoscopy, but with longer intubation times²⁶.

Airtraq

Airtraq is a rigid disposable laryngoscope with two channels: one for the ETT and the other for a cold-light source with anti-



Fig. 4 – Storz videolaryngoscope: insertion and view of the vocal chords

Source: Authors.

fog lenses, prisms, and mirrors that transmit the image to a screen located on the opposite end or to an external Bluetooth connected monitor (Fig. 5)⁷.

Different sizes are available that allow for tube diameters ranging from 2.5 mm up to 8.5 mm and has presentations for nasal intubation. 35–37F double-lumen tubes have also been inserted with the Airtraq⁷.

Studies have shown that the learning curve for personnel trained in laryngoscopy is shorter, with a shorter intubation time, greater number of successful intubations, and less external larynx manipulation. In patients with manual alignment of the cervical spine, the Airtraq requires less vertebral movement as demonstrated in radiological studies^{27,28}.

The limitations for its use are: an oral aperture of at least 20 mm, a reduced thyromental distance, blood or secretions in the airway, and tearing of the balloon cuff due to insufficient lubrication in the tube channel²⁷.

Mcgrath videolaryngoscope

The McGrath videolaryngoscope consists of a blade called a CameraStick, whose length can be modified for use in children



Fig. 5 – AirTraq videolaryngoscope.

Source: www.airwayskills.co.nz with permission.

and adults. It has a source of LED light and a video camera in its distal tip. A disposable blade covers the CameraStick and can be used as a lever in the glottic cavity. Attached to the handle is a 2.5" LCD screen whose angle can be changed⁷.

There are reports that the McGrath videolaryngoscope can convert a Cormack-Lehane grade 3 or 4 glottis into a 1 or 2, with success rates of up to 95%²⁹.

Other studies have shown that, although the glottis visualization improves with the use of the McGrath, the time required, the number of failed intubations, and external manipulations needed were greater compared to the Macintosh laryngoscope³⁰.

There have been descriptions of lesions in the oral cavity produced when the observer moves the tube without visualizing the structures through which it is passing through and it implies a learning curve to be used in cases of difficult airway³⁰.

Bullard laryngoscope

The Bullard rigid fiber optic laryngoscope can be introduced into patient's mouth with a minimum oral aperture of 6 mm so that a glottic view can be achieved without hyperextension of the cervical spine. It consists of a metallic curved L-shaped blade of which the distal tip can be attached to a plastic piece to make it longer for intubations in large patients.

A light source, the optical lens, and a channel for aspiration or oxygen flow emerge from the posterior part of the blade. This allows for the attachment of the video camera and has different presentations for newborns and pediatric patients^{31,32}.

The Bullard laryngoscope does not have a panoramic view, and if it is not introduced along the midline, the vocal cords may not be seen. In simulated settings of cervical trauma



Fig. 6 – McGrath Videolaryngoscope.

Source: www.aircraftmedical.com with permission.

and rapid sequence intubation, it is effective in securing the intubation, but with prolonged times³³. Compared to laryngeal mask intubation, it shows a non-significant tendency of greater effectiveness in intubation with an aligned cervical spine³⁴. As with other devices, it involves a learning curve and the recommended setting is with non-urgent airways³⁵. It can be used for nasal intubation (Figs. 6 and 7).

Discussion

Difficult airway management continues to be a challenge in the practice of anesthesia. The identification of a Cormack-Lehane grade 3 or 4 does not closely correlate to the difficulty of intubation since the majority of these patients can be successfully intubated with the help of a stylet or a bougie^{36,37}. However, the optimization of larynx visualization with devices like videolaryngoscopes does not guarantee successful intubation. The efficacy of videolaryngoscopes in cases of difficulty airway has limited evidence. Devices like the Airtraq, the Glidescope, and the Bullard have been recommended in cases where difficult airways and failed intubation occurs with direct laryngoscopy, but in experienced hands³⁸⁻⁴⁰. It has been



Fig. 7 – Bullard laryngoscope.

Source: www.airwayskills.co.nz with permission.

shown that success with videolaryngoscopes is related to experience in management, with a learning curve that generally does not exceed 10 patients.

Studies that compare the different types of videolaryngoscopes to conventional laryngoscopy have so far shown an improvement in the visualization of structures to approximately grades 1 or 2. However, there continues to be controversy in the duration and success rate of the intubation^{36,39,41}. In inexperienced hands in the emergency setting, the use of videolaryngoscopes has been shown to increase the success rate for intubation after the first attempt⁴²⁻⁴⁵.

In obese patients, in which intubation can be difficult in up to 15% of patients⁴⁶, videolaryngoscopes have shown improvement in the visualization of the larynx with no difference found in intubation times^{47,48}. In a large percentage of patients (3.7%)⁴⁹, despite the good visualization, intubation was not possible.

Up to the present, counterindications for the use of videolaryngoscopes have not been described, and the associated complications, such as lesions in the airway, are only starting to be described⁴⁹. Altered anatomy has been mentioned as a possible predictor of failure with the Glidescope²⁰ and advancing the tube without visualizing the structures through which it is passing can be an important cause of complications. Likewise, there are no studies published in terms of cost analysis that compare videolaryngoscopes with direct laryngoscopes. Nevertheless, there are publications that compare disposable and reusable videolaryngoscopes, showing similar values⁵⁰. Apparently the net cost of a videolaryngoscopy is higher than that of a direct laryngoscope in terms of its price, maintenance, battery, hygiene, training, etc. That said, more cost-effectiveness studies are needed to support this theory. Compared to fibrobronchoscopy, they are less expensive, but the evidence of their efficacy in difficult airway situations continues to be weak. Therefore, fibrobronchoscopy continues to be the gold standard.

Conclusion

Although recently the use of videolaryngoscopes has been mentioned in the algorithm for difficult airway

management^{4,51} with type A evidence of improvement in laryngeal visualization, controversies continue to persist with regard to the value of this device in the management of difficult airway and safety for patients. Therefore, it continues to be a topic of investigation⁵¹. For the time being, intubation with fibrobronchoscopy in conscious patients continues to be the safest method of managing an anticipated difficult airway¹.

Funding

None.

Conflicts of interest

The authors have no conflicts of interest to declare.

REFERENCES

1. Apfelbaum JL, Hagberg CA, Caplan RA, Blitt CD, Connis RT, Nickinovich DG, et al. Practice guidelines for management of the difficult airway: an updated report by the American Society of Anesthesiologists task force on management of the difficult airway. *Anesthesiology*. 2013;118:251–70.
2. American Society of Anesthesiologists Task Force on Management of the Difficult Airway. Practice guidelines for management of the difficult airway: an updated report by the American Society of Anesthesiologists Task Force on Management of the Difficult Airway. *Anesthesiology*. 2003;98:1269–77 [Erratum in: *Anesthesiology* 2004;101(2):565].
3. Henderson JJ, Popat MT, Latto IP, Pearce AC. Difficult Airway Society guidelines for management of the unanticipated difficult intubation. *Anaesthesia*. 2004;59:675–94.
4. Crosby ET, Cooper RM, Douglas MJ, Doyle DJ, Hung OR, Labrecque P, et al. The unanticipated difficult airway with recommendations for management. *Can J Anaesth*. 1998;45:757–76.
5. Benumof JL. Management of the difficult adult airway with special emphasis on awake tracheal intubation. *Anesthesiology*. 1991;75:1087–110 [Review. Erratum in: *Anesthesiology*, 1993;78:224].
6. Guzmán J. Videolaringoscopios. *Rev Chil Anest*. 2009;38:135–44.
7. Niforopoulou P, Pantazopoulos I, Demestiha T, Koudouna E, Xanthos T. Video-laryngoscopes in the adult airway management: a topical review of the literature. *Acta Anaesthesiol Scand*. 2010;54:1050–61.
8. Asai T. Videolaryngoscopes: do they truly have roles in difficult airways? *Anesthesiology*. 2012;116:515–7.
9. Pott LM, Murray WB. Review of video laryngoscopy and rigid fiberoptic laryngoscopy. *Curr Opin Anaesthesiol*. 2008;21:750–8.
10. Henderson JJ, Suzuki A. Rigid indirect laryngoscope insertion techniques. *Anaesthesia*. 2008;63:323–4.
11. Greenland KB, Segal R, Acott C, Edwards MJ, Teoh WH, Bradley WP. Observations on the assessment and optimal use of videolaryngoscopes. *Anaesth Intensive Care*. 2012;40:622–30.
12. Leviatan RM, Heitz JW, Sweeney M, Cooper RM. The complexities of tracheal intubation with direct laryngoscopy and alternative intubation devices. *Ann Emerg Med*. 2011;57:240–7.
13. Rai MR, Dering A, Vergheze C. The Glidescope system: a clinical assessment of performance. *Anaesthesia*. 2005;60:60–4.
14. Sakles JC, Rodgers R, Keim SM. Optical and video laryngoscopes for emergency airway management. *Intern Emerg Med*. 2008;3:139–43.
15. Jones PM, Harle CC, Turkstra TP. The GlideScope Cobalt videolaryngoscope – a novel single-use device. *Can J Anaesth*. 2007;54:677–8.
16. Nakstad AR, Sandberg M. The GlideScope Ranger video laryngoscope can be useful in airway management of entrapped patients. *Acta Anaesthesiol Scand*. 2009;53:1257–61.
17. Stroumpoulis K, Pagoulatou A, Violari M, Ikonomou I, Kalantzi N, Kastrinaki K, et al. Videolaryngoscopy in the management of the difficult airway: a comparison with the Macintosh blade. *Eur J Anaesthesiol*. 2009;26:218–22.
18. Lai HY, Chen IH, Hwang FY, Lee Y. The use of the GlideScope for tracheal intubation in patients with ankylosing spondylitis. *Br J Anaesth*. 2006;97:419–22.
19. Griesdale DE, Liu D, Mckinney J, Choi PT. Glidescope video-laryngoscopy versus direct laryngoscopy for endotracheal intubation: a systematic review and meta-analysis. *Can J Anaesth*. 2012;59:41–52.
20. Aziz MF, Healy D, Kheterpal S, Fu RF, Dilman D, Brambrink M. Routine clinical practice effectiveness of the Glidescope in difficult airway management: an analysis of 2,004 Glidescope intubations, complications, and failures from two institutions. *Anesthesiology*. 2011;33:40–6.
21. Airway management products by kingsystems [Internet homepage]. USA: King Systems; 2010. Available from: <http://www.owntheairway.com> [updated 2014, consulted 2014 March].
22. Akihisa Y, Maruyama K, Koyama Y, Yamada R, Ogura A, Andoh T. Comparison of intubation performance between the King Vision and Macintosh laryngoscopes in novice personnel: a randomized, crossover manikin study. *J Anesth*. 2014;28:51–7.
23. Murphy LD, Kovacs GJ, Reardon PM, Law JA. Comparison of the king vision video laryngoscope with the macintosh laryngoscope. *J Emerg Med*. 2014;14:1–8.
24. Kaplan MB, Ward DS, Berci G. A new video laryngoscope – an aid to intubation and teaching. *J Clin Anesth*. 2002;14:620–6.
25. Kaplan MB, Hagberg CA, Wards DS, Brambrink A, Chhibber AK, Heidegger T, et al. Comparison of direct and video-assisted views of the larynx during routine intubation. *J Clin Anesth*. 2006;18:357–62.
26. Aziz MF, Dillman D, Fu R, Brambrink AM. Comparative effectiveness of the C-MAC video laryngoscope versus direct laryngoscopy in the setting of the predicted difficult airway. *Anesthesiology*. 2012;116:629–33.
27. Castañeda PM, Batllori M, Gomez-Ayechu M, Iza J, Unzué P, Martín MP. Airtraq optical laryngoscopy. *An Sist Sanit Navar*. 2009;32:75–83.
28. Amathieu R, Combes X, Abdi W, Housseini LE, Ezzoug A, Dinca A, et al. An algorithm for difficult airway management, modified for modern optical devices (Airtraq laryngoscope; LMA CTrach™): a 2-year prospective validation in patients for elective abdominal, gynecologic, and thyroid surgery. *Anesthesiology*. 2011;114:25–33.
29. Noppens RR, Mobus S, Heid F, Schimidtmann I, Werner C, Piepho T. Evaluation of the McGrath Series 5 videolaryngoscope after failed direct laryngoscopy. *Anaesthesia*. 2010;65:716–20.
30. Frohlich S, Borovickova L, Foley E, O'Sullivan E. A comparison of tracheal intubation using the McGrath or the Macintosh laryngoscopes in routine airway management. *Eur J Anaesthesiol*. 2011;28:465–7.

31. Harberg C. *Benumof's Airway Management*. 2nd ed. Philadelphia: MOSBY ELSEVIER; 2007, ISBN 978-0-323-02233-0. p. 558.
32. The Virtual Airway Device Intubation techniques and tutorials [Internet homepage]. Florida: University of Florida; 2006. Available from: <http://vam.anest.ufl.edu/airwaydevice/bullard/index.html> [updated 2011, consulted 2013].
33. Watts AD, Gelb AW, Bach DB, Pelz DM. Comparison of the Bullard and Macintosh laryngoscopes for endotracheal intubation of patients with a potential cervical spine injury. *Anesthesiology*. 1997;87:1335-1342.
34. Nileshwar A, Thudamalinne A. Comparison of intubating laryngeal mask airway and Bullard laryngoscope for oro-tracheal intubation in adult patients with simulated limitation of cervical movements. *Br J Anaesth*. 2007;99:292-6.
35. MacQuarrie K, Hung OR, Law JA. Tracheal intubation using Bullard laryngoscope for patients with a simulated difficult airway. *Can J Anaesth*. 1999;46:760-5.
36. Cook TM. A new practical classification of laryngeal view. *Anaesthesia*. 2000;55:274-9.
37. Xue FS, Liao X, Liu JH. Comparative performance of direct and video laryngoscopes in patients with predicted difficult airway. *Anesthesiology*. 2012;117:911-25.
38. Mihai R, Blair E, Kay H, Cook TM. A quantitative review and meta-analysis of performance of non-standard laryngoscopes and rigid fiberoptic intubation aids. *Anaesthesia*. 2008;63:745-60.
39. Galán JC, Charco P, Sadarangani A. Inducción inhalatoria y anestesia tópica de la vía aérea guiada por videolaringoscopia C-MAC pala d-Blade en un paciente con miastenia gravis y vía aérea difícil. *Rev Colomb Anestesiología*. 2013;41:287-90.
40. Healy DW, Maties O, Hovord D, Kheterpal S. A systematic review of the role of videolaryngoscopy in successful orotracheal intubation. *BMC Anesthesiol*. 2012;12:32.
41. Corso RM, Piraccini E, Agnoletti, Gambale G. Comparison of video laryngoscopes with direct laryngoscopy for tracheal intubation: some clarification needed. *Eur J Anaesthesiol*. 2012;29:495-8.
42. Grisaldale DE, Liu D, McKinney J, Choi PT. Glidescope® video-laryngoscopy versus direct laryngoscopy for endotracheal intubation: a systematic review and meta-analysis. *Can J Anaesth*. 2012;59:41-52.
43. Kory P, Guevarra K, Mathew JP, Hegde A, Mayo PH. The impact of video laryngoscopy use during urgent endotracheal intubation in the critically ill. *Anesth Analg*. 2013;117:144-9.
44. Maruyama K, Yamada T, Kawakami R, Kamata T, Yokochi M, Hara K. Upper cervical spine movement during intubation: fluoroscopic comparison of the AirWay Scope McCoy laryngoscope, and Macintosh laryngoscope. *Br J Anaesth*. 2008;100:120-4.
45. Aziz M. Use of video-assisted intubation devices in the management of patients with trauma. *Anesthesiol Clin*. 2013;31:157-66.
46. Juvin P, Lavaut E, Dupont H, Demetriou M, Dumoulin JL, Desmont JM. Difficult tracheal intubation is more common in obese than in lean patients. *Anesth Analg*. 2003;97:595-600.
47. Bathory I, Granges JC, Frascarolo P, Magnusson L. Evaluation of the Video Intubation Unit in morbid obese patients. *Acta Anaesthesiol Scand*. 2010;54:55-8.
48. Marrel J, Blanc C, Frascarolo P, Magnusson L. Videolaryngoscopy improves intubation condition in morbidly obese patients. *Eur J Anaesthesiol*. 2007;24:1045-9.
49. Cooper RM, Pacey JA, Bishop MJ, McCluskey SA. Early clinical experience with a new videolaryngoscope (GlideScope) in 728 patients. *Can J Anaesth*. 2005;52:191-8.
50. Tvede MF, Kristensen MS, Nyhus-Andreasen M. A cost analysis of reusable and disposable flexible optical scopes for intubation. *Acta Anaesthesiol Scand*. 2012;56:577-84.
51. Paolini JB, Donati F, Drolet P. Review article: video-laryngoscopy: another tool for difficult intubation or a new paradigm in airway management? *J Can Anesth*. 2013;60:184-91.